

Lab Number 334835

TABLE 1. Results of Particle Filtration Test  
Sample ID: NT-Viral Mask - 5126

SAMPLE NUMBER	AVERAGE SAMPLE COUNTS	AVERAGE CONTROL COUNTS	FILTRATION EFFICIENCY
1	45	13188	99.66%
2	51	13676	99.63%
3	62	14157	99.56%
4	49	10955	99.55%
5	58	13088	99.56%

SAMPLE AREA TESTED: 91.5 cm<sup>2</sup>

PARTICLE SIZE: 0.1 μm (0.102 ± 0.003 μm)

PARTICLE BACKGROUND: <1 particles/min

AVERAGE FILTRATION EFFICIENCY: 99.59%

STANDARD DEVIATION: 0.048